

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 07 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130187
Name		DAN PASQUARIELLO	ID# 2620 Date 5-26-10
A	Agency	DILLINGHAM POLICE DEPT.	Phone # 842-5354
Instrument Location		DILLINGHAM POLICE DEPT.	
Alco S/N		78853	Target Value .083 High Pressure 350 psi
B	Alco Test Values		1st Alco .082 2nd Alco .082
Signature		COB 6/10/10	
(OVER)			



I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 24th day of June, 2010.

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130187

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130187

MAY 26, 2010

OPERATOR'S NAME:

PASQUARIELLO, DAN

OPERATOR'S NUMBER: 2620

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: 1234567890

DEPT/AGENCY: DL61

CASE/REPORT: 1234567890

TEST TYPE: U

ALCO TARGET VALUE: .083

ALCO S/N: 78853

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 30.17 in

ALCO TARGET	.083	18:39
BLANK TEST	.000	18:40
INTERNAL STANDARD	VERIFIED	18:40
ALCO TV 30.17 in	.082	18:41
BLANK TEST	.000	18:42
SUBJECT SAMPLE	.000	18:43
BLANK TEST	.000	18:44
ALCO TV 30.17 in	.082	18:44
BLANK TEST	.000	18:45

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130187

MAY 26, 2010

TIME 18:46

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 43c

BAROMETER: 30.17 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~